



Reflecting the Truth

# CERTIFICATE OF CALIBRATION

Issued By :

**YADAV MEASUREMENTS PVT. LTD.**

PLOT No. 19-20, HARIDAS JI KI MAGRI

TRIDENT ROAD, UDAIPUR (RAJ.), INDIA 313004

Email # yadav.measurements@ymflabs.com

Phone No. : 0294-2434050 Telefax : 0294-2434067



C-035  
Electro- Technical



0616

Certificate number:-YMPL/91675/11415

Page 1 to 2

1. Name and address of client :-

SONAA ENGINEERS PVT. LTD.  
308-310 SUBHAM COMPLEX  
NEW FETEHPURA, UDAIPUR (RAJ.)

2. Reference:

Service request form number: 2006-2007/1609

Date: 17<sup>th</sup> February, 2007

Date of the receipt of the equipment: 17<sup>th</sup> February, 2007

Condition on item: Satisfactory

3. Calibration certificate:

Date of issue: 22<sup>nd</sup> February, 2007

Date of calibration: 17<sup>th</sup> February, 2007

Calibration due date: 17<sup>th</sup> February, 2010

4. Description of equipment under calibration: -

Name: ACA LEAKAGE TESTER

Serial number: C53471

Model: DL-650

Make : LUTRON

5. Environmental conditions of measurements:

Temperature : 25±2°C

Relative humidity: ≤70%

6. Description of reference standards used :

Reference standard	Calibration valid upto	Traceability	Parameters
Three phase precision measuring instrument (TPZ-303) IS 1055 with precision CT IS 814	27 <sup>th</sup> September, 2007	DKD Germany via YMPL Standards	a.c. current

### Remarks :-

1. For k=2

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS/NABL requirements.

2. For k=2

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k, which for a t-distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

3. This report is specific for E.U.C. for environmental and other conditions mentioned in this report.

4. The reported uncertainty applies only to the measured value and gives no indication of the long term stability of the device.

5. The instrument has been calibrated for calibration points required by customer.

6. The calibration due date has been mentioned as requested by customer in writing.

This Certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service and the NABL (Govt. of India). It provides traceability of measurement to recognized national standards, and to units of measurement realized at the National Physical Laboratory or other recognized national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory

Checked by :

*RNP*

*RSD*

Approved by :

*Vinod Talwar*

Sign :

Name :

*APL*

# CERTIFICATE OF CALIBRATION

UKAS Accredited calibration laboratory No. 0616  
NABL Accredited calibration laboratory No. C-035

Issued by :  
Yadav Measurements Pvt. Ltd., Udaipur

Certificate number: YMPL/91675/11415

Page 2 of 2

7. Calibration procedure :  
Equipment under calibration was calibrated for a.c. current measurement, electronic source VCS-320 with C.B.S. setup was used and fed to reference TPZ-303 and EUC. At least 6 measurements were made for each value and an average was reported. EUC was powered up from its internal supply source (i.e. 9 Volt battery) and warmed up for at least 15 Minute.
8. Results: The Digital clamp multimeter has been calibrated for a.c. current measurement. The expanded uncertainty of our measurement is given in front of each measurement & calculated at 95 % CL. The results are as following.

## 9. Calibration results:-

Parameter	Range	Measured value	Standard value	± Expanded uncertainty (%)	Coverage factor(k)
A.C. current: (50Hz, sine wave)	200mA	19.7mA	20.05mA	0.27	2.00 *
		99.5mA	100.42mA	0.06	2.00
		178.5mA	180.64mA	0.04	2.00
	20A	0.98A	1.001A	0.53	2.00 *
		4.99A	5.010A	0.12	2.00
		9.88A	9.916A	0.07	2.00
		14.84A	15.015A	0.05	2.00
	200A	9.8A	9.92A	0.53	2.00 *
		25.1A	25.02A	0.23	2.00
		50.0A	50.04A	0.12	2.00
		98.8A	99.26A	0.07	2.00
		118.0A	118.97A	0.06	2.00

## REMARKS:

1. "\*" indicates that the uncertainty quoted is dominated by the uncertainty due to resolution of the instrument being calibrated, for which a rectangular probability distribution has been assumed.

Checked by :

RNP

RSD

Approved by :

DPZ