

 <p>Reflecting the Truth</p>	<p align="center">CERTIFICATE OF CALIBRATION</p> <p>Issued By : YADAV MEASUREMENTS PVT. LTD. PLOT No. 19-20, HARIDAS JI KI MAGRI TRIDENT ROAD, UDAIPUR (RAJ.), INDIA 313004 Email # yadav.measurements@ymllabs.com Phone No. : 0294-2434050 Telefax : 0294-2434067</p>	 <p align="center">C-035 Electro- Technical</p>	 <p align="center">UKAS CALIBRATION 0616</p>
---------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------

Certificate number: YMPL/91671/11413

PAGE 1 OF 2

1. Name & Address of Client:-

SONAA ENGINEERS PVT. LTD.
308-310 SUBHAM COMPLEX
NEW FETEHPURA, UDAIPUR (RAJ.)

2. Reference:-

Service request form number : 2006-2007/1609
Date: 17th February, 2007
Date of receipt item : 17th February, 2007
Condition of item on receipt : Satisfactory

3. Calibration detail :-

Date of issue: 22nd February, 2007
Date of calibration : 17th February, 2007
Due date of calibration: 17th February, 2010

4. Description of equipment under calibration:-

Name: Digital ammeter with testing set
Serial number : 0314125 with 9606-107A
Model : SMP45RS

5 Environmental conditions of measurements:

Temperature: 25±2°C
Relative humidity: <70%

6 Description of reference standards used :

Reference standard	Calibration valid upto	Traceability	Parameters
Three phase reference meter I/S 1055 with Three phase precision C.T.(I/S 814)	27 th September,2007	DKD Germany via YMPL	AC CURRENT

Remarks :-

- For k=2
The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS/NABL requirements.
- For k=2
The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k, which for a t-distribution corresponds to a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.
- This report is specific for E.U.C. for environmental and other conditions mentioned in this report.
- The reported uncertainty applies only to the measured value and gives no indication of the long term stability of the device.
- The instrument has been calibrated for calibration points required by customer.
- The calibration due date has been mentioned as requested by customer in writing.

This Certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service and the NABL (Govt. of India) It provides traceability of measurement to recognized national standards, and to units of measurement realized at the National Physical Laboratory or other recognized national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

<p>Checked by : <u>RMP</u> <u>RSD</u></p>	<p>Approved by : <u>Dinesh Palival</u> Sign : <u>[Signature]</u> Name : <u>[Signature]</u></p>
-------------------------------------------	----------------------------------------------------------------------------------------------------------

Certificate number: YMPU/91671/11413

PAGE 2 OF 2

7. Calibration procedure:

Equipment under calibration (having build in CT & current source) has been calibrated for a.c. current. The instrument was supplied from ac mains of 240V, 50Hz from UPS. The ac current output from the EUC was connect in series with reference measuring CT. The secondary of CT is connected to the reference three phase precision measuring instrument for standard values. No other load is connected to EUC during calibration .the measured values reported were reported from ammeter connected on the EUC. At least 6 measurements have been mode for each value and an average value has been reported. EUC is warmed up for at least half an hour.




8 Results:- The DPM Ammeter has been calibrated for a.c. current ranges. The Expanded uncertainty of our measurements is given in front of each measurement & calculated at 95% CL. The results are given bellow.

8. Calibration results:

Parameter	Range	Measured value (Amp.)	Standard value (Amp.)	± Expanded Uncertainty (%)	Coverage Factor (k)
a.c. current (50Hz sine wave)	50A,50Hz	48.44	49.373	0.03	2.00
		39.28	40.228	0.12	2.00
		29.63	30.290	0.08	2.00
		19.92	20.408	0.05	2.00
		9.97	10.127	0.19	2.00
		4.93	5.0623	0.49	2.00
		1.97	2.0245	0.22	2.00
		0.96	1.0013	0.30	2.00
		0.48	497.86	1.07	2.00 *

Remarks :

1. "*"Indicates that "the uncertainty quoted is dominated by resolution of EUC for which rectangular probability distribution has been assumed.

Checked by : <div style="text-align: center; font-family: cursive;">   </div>	Approved by : <div style="text-align: center; font-family: cursive;">  </div>
--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------